

# EC Declaration of Conformity

We Trumeter  
of Pilot Mill  
Alfred Street  
Bury, Lancs  
England  
BL9 9JR

*in accordance with the following Directives:*

2014/35/EU	The Low Voltage Directive (LVD)
2014/30/EU	The Electromagnetic Compatibility Directive (EMC)
2011/65/EU	The Restriction of Hazardous Substances Directive (RoHS 2)

*hereby declare under our sole responsibility that:*

Equipment	Hour Meter
Model number(s)	711-0109, 711-0113, 711-0114, 711-0130, 711-0160, 711-0162, 711-0163, 711-0180, 711-0190, 711-0217, 711-0013, 711-0019, 711-0020, 711-0041, 711-0120, 711-0131, 711-0138, 711-0144, 711-0150, 711-0161, 711-0164, 711-0182, 711-0187, 711-0191, 711-0192, 711-0195, 711-0100

*are in conformity with the applicable requirements of the following documents:*

<u>Reference</u>	<u>Title</u>	<u>Edition (Date)</u>
IEC 61010-1	Safety requirements for electrical equipment for measurement, control, and laboratory use	3.0 (2010-06)
IEC 61326-1	EMC requirements for equipment designed for measurement, control and laboratory use	2.0 (2012-07)
IEC 61000-4-2	Electrostatic discharge (ESD) immunity test	2.0 (2008-12)
IEC 61000-4-3	Radiated, radio-frequency, electromagnetic field immunity test	3.2 (2010-04)
IEC 61000-4-4	Electrical fast transient/burst immunity test	3.0 (2012-04)
IEC 61000-4-5	Surge immunity test	3.0 (2014-05)
IEC 61000-4-6	Immunity to conducted disturbances, induced by radio-frequency fields	4.0 (2013-10)
IEC 61000-4-8	Power frequency magnetic field immunity tests	2.0 (2009-09)
IEC 61000-4-11	Immunity to voltage dips, interruptions and variations.	2.0 (2004-03)
CISPR-11	Conducted and Radiated Emissions	6.0 (2016-06)

I hereby declare that the equipment named above has been designed to comply with the relevant sections of the above referenced specifications. The unit complies with all applicable Essential Requirements of the Directives.

Signed by: .....

Name:  Gill Copitch

Position: CEO

Done at Bury

On 13/05/2025



Document number: DOC5737

Issue 4

**Trumeter Technologies Ltd**  
Pilot Mill  
Alfred Street  
Bury BL9 9EF

tel +44 (0)161 674 0960  
fax +44 (0)161 480 8773  
email [info@trumeter.com](mailto:info@trumeter.com)  
web [www.trumeter.com](http://www.trumeter.com)

Registration number: 7115948