

EC Declaration of Conformity

We Trumeter
of Pilot Mill
Alfred Street
Bury, Lancs
England
BL9 9EF

in accordance with the following Directives:

2014/35/EU	The Low Voltage Directive (LVD)
2014/30/EU	The Electromagnetic Compatibility Directive (EMC)
2011/65/EU	The Restriction of Hazardous Substances Directive (RoHS 2)

hereby declare under our sole responsibility that:

Equipment	LCD Digital Panel Meter
Model number(s)	DPM951-T, DPM952-T

are in conformity with the applicable requirements of the following documents:

<u>Reference</u>	<u>Title</u>	<u>Edition (Date)</u>
IEC 61010-1	Safety requirements for electrical equipment for measurement, control, and laboratory use	3.0 (2010-06)
IEC 61326-1	EMC requirements for equipment designed for measurement, control and laboratory use	2.0 (2012-07)
IEC 61000-4-2	Electrostatic discharge (ESD) immunity test	2.0 (2008-12)
IEC 61000-4-3	Radiated, radio-frequency, electromagnetic field immunity test	3.2 (2010-04)
IEC 61000-4-4	Electrical fast transient/burst immunity test	3.0 (2012-04)
IEC 61000-4-6	Immunity to conducted disturbances, induced by radio-frequency fields	4.0 (2013-10)
CISPR-11	Conducted and Radiated Emissions	6.0 (2016-06)

I hereby declare that the equipment named above has been designed to comply with the relevant sections of the above referenced specifications. The unit complies with all applicable Essential Requirements of the Directives.

Signed by:

Name:  Gil Copitch
Position: Director of Product Development
Done at Bury
On 10/05/2018



Document number:
(issue 1) DOC5718

Trumeter Technologies Ltd	tel	+44 (0)161 674 0960
Pilot Mill	fax	+44 (0)161 480 8773
Alfred Street	email	info@trumeter.com
Bury BL9 9EF	web	www.trumeter.com

Registration number: 7115948